



# VERIFICATION

No. 0B240909.NHLCT12

Test Report / Technical Construction File no. LIFE242105PD

Verification's  
Holder:

Ningbo Haishu Life Medical Technology Co., Ltd.  
No.1 Jinghui Road, Hengjie Town 315181 Haishu, Ningbo,  
Zhejiang, People's Republic Of China

Verification  
ECM Mark:



Product:

Chamber of Steam Sterilizers

Model(s):

LF08L, LF12L, LF18L, LF22L, LF23L

(see the following annex)

Verification to:

Standard:

---

related to CE Directive(s):

2014/68/EU (Pressure Equipment) Annex II

Module: A

**This document has been issued in accordance with the European Commission's note of 14 September 2022 ref. Ares (2022) 6342894 concerning voluntary certifications with non-notified procedure and for products intended for third countries outside the EU market.**

The manufacturer has voluntarily decided to submit its documents concerning the above-mentioned product for verification. Ente Certificazione Macchine confirms that the documentation made available and immediately returned to it, as containing sensitive data, meets the essential requirements of the above-mentioned directives. The verification activity carried out exclusively concerned the technical documentation and no verification was carried out on the product. This document cannot replace the EC Declaration of Conformity. The above conformity mark can be affixed to the technical documentation in accordance with the ECM regulation on its issue and use, published on the website [www.entecerma.it](http://www.entecerma.it)

**Issuance date: 09 September 2024**

**Expiry date: 08 September 2029**

For online check:



Approver  
Ente Certificazione Macchine  
Legal Representative  
Luca Bedonni



**Ente Certificazione Macchine Srl**

Via Ca' Bella, 243 – Loc. Castello di Serravalle – 40053 Valsamoggia (BO) - ITALY  
☎ +39 051 6705141 📠 +39 051 6705156 ✉ info@entecerma.it 🌐 [www.entecerma.it](http://www.entecerma.it)

# Annex I



No. 0B240909.NHLCT12

Test Report / Technical Construction File no. LIFE242105PD

Max. Allowable Pressure  $P_s$  (bar): 2.7  
 Nominal size Volume  $V$  (l): 8, 12, 18, 22, 23  
 Max. Allowable temperature  $T_s$  (°C): 140  
 Wall thickness (mm): 2  
 Medium: Steam

